

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 13 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130183 ✓
Name	HUTT, William T.	ID#	3359 Date 07/09/09 ✓
A Agency	Homer Police Dept. / Jail	Phone #	235-3150
Instrument Location	4060 Heath St. Homer, AK 99603 ✓		
B Alco S/N	X 301599 ✓	Target Value	.086 ✓
		High Pressure	925psi
Alco Test Values	.086 ✓	.086 ✓	
	1 st Alco	2 nd Alco	
Signature	William T. Hunt		
083 7/14/09			
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Jul, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

JULY 09, 2009

OPERATOR'S NAME: HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
ETOH

O.L. #: 01234567
DEPT/AGENCY: HOM1
CASE/REPORT: 09-0000
TEST TYPE: U
ALCO TARGET VALUE: .006
ALCO S/N: X301599

— BREATH ANALYSIS —

.006 ADJUSTED FOR 30.01 in
ALCO TARGET .006 03:10
BLANK TEST .000 03:11
INTERNAL STANDARD VERIFIED 03:11
ALCO TV 30.01 in .006 03:11
BLANK TEST .000 03:12
SUBJECT SAMPLE .000 03:13
BLANK TEST .000 03:14
ALCO TV 30.01 in .006 03:14
BLANK TEST .000 03:15

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

JULY 09, 2009
TIME 03:06

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.01 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@AB CDEF6
HIJKL MNOPQRSTU VWXYZ[\]^_`abcde fghij klmno
pqrstu vwxyz{|}~■